

Appendix II

Name of the Item & specifications	Quantity	Rate
UV-visible spectrophotometer with Double beam optics direct ratio system with Dynode feedback, Wavelength Range: 900 nm to 190 nm, \pm 0.3nm or more accuracy, Abs accuracy, Spectrum, Photometric for multiple wavelengths, quantization, Kinetics measurements are required for data analysis, nano volume assay is preferable, should be supplied with PC, suitable printer and software	01 Unit	

Name of the Tenderer:

Address:

Phone and Fax Number:

D.D. No., Date & Amount

Signature with stamp of the authorized dealer